

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>0390</b></p> <p>Accredited to <b>ISO/IEC 17025:2005</b></p>	<p><b>Microprise Ltd</b></p> <p><b>Issue No: 015    Issue date: 21 February 2018</b></p>	
	<p><b>Unit 3, Plot 7F</b> <b>Claymore</b> <b>Tame Valley Commerce Centre</b> <b>Tamworth</b> <b>Staffordshire</b> <b>B77 5DQ</b></p>	<p><b>Contact: Mr G Margetts</b> <b>Tel: +44 (0)1827 261554</b> <b>Fax: +44 (0)1827 261552</b> <b>E-Mail: graham@microprise.co.uk</b> <b>Website: www.microprise.co.uk</b></p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k = 2)	Remarks
<p>RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED</p>			
LENGTH			
Plain plug gauges (parallel)	1 to 50 diameter 50 to 100 diameter 100 to 150 diameter 150 to 200 diameter 200 to 300 diameter	1.2 1.21.5 2.0 3.0	
Plain plug gauges (taper)			
Taper up to 1 in 8 on diameter	3 to 50 diameter 50 to 100 diameter 100 to 200 diameter 200 to 300 diameter	3.0 4.0 5.0 6.0	] on diameter
Taper above 1 in 8 And up to 1 in 3	3 to 50 diameter 50 to 100 diameter 100 to 200 diameter 200 to 300 diameter	5.0 6.0 7.0 8.0	
Plain ring gauges (parallel)	10to 50 diameter 50 to 100 diameter 100 to 150 diameter	1.8 2.0 2.5	
Length gauges, flat and spherical ended (excluding length bars)	0 to 300	1.0 + (8.0 x length in m)	
Plain gap gauges (parallel)	0.5 to 100 100 to 200 200 to 300	3.0 5.0 8.0	
Receiver, position and profile gauges, jigs, fixtures	0 to 400 x 200 x 200	Dependant on size and features Minimum per co-ordinate: 3.0 + 10.0 x length in m	
<p>END</p>			